

VERIFICATION OF CONFORMITY

Certificate No.: B-S2506D2547

Client ID: CD1872

Beide
Compliance Laboratory**Holder**.....: Suzhou Jiujiuba Medical Equipment Co., Ltd**Address**.....: No. 86, Renmin Road, Zhaofeng, Leyu Town, Zhangjiagang City**Manufacturer**.....: Suzhou Jiujiuba Medical Equipment Co., Ltd**Address**.....: No. 86, Renmin Road, Zhaofeng, Leyu Town, Zhangjiagang City**Product**.....: Electric climbing wheelchair**Model No**.....: WD58-8**Technical Data**.....: Battery Charger:
Input: 100-240V~, 50/60Hz, 2.5A, CLASS II
Output: DC 29.4V, 2.8A
Battery: 25.5V, 12.5Ah**Verification to**.....: EN 12184:2022

The test results apply only to the particular sample tested and to the specific tests carried out. Technical Report and documentation are at the Holder's disposal.

This certificate applies specifically to the sample investigated in our test reference number only. This is to certify that the tested sample is in conformity with above listed standard or required test method.

Date of issue: 2025-07-22**Expiry date:** 2028-07-21

Certification Manager

**Beide (Shenzhen) Product Service Limited**

Address: 6F, Bldg E, Hourui 3rd Ind Zone, Xixiang, Bao'an Dist, Shenzhen, China

E-mail: admin@beide.com Phone: +86 0755 27454498 www.beide.com

